

QUALITY CERTIFICATE
СЕРТИФИКАТ КАЧЕСТВА

Essential requirements in accordance with the Directive DIR 98/79/EC
Необходимые требования в соответствии с директивой 98/79/EC

The manufacturer / Наименование производителя Foracare Suisse AG. & TaiDoc Technology Corporation

Product / Продукция:

Gamma DM Blood Glucose Test Strip Lot no. / Тест полоски C/H

TD23B301-BOR

Quantity / Количество: 8000

Manufacture date / Дата изготовления 2023/02/01

Strip expiration date / Дата годности тест полосок до 2024/11/01

The final test for accuracy / Заключительный тест на точность:

Requirements / Требования	Result / Результат	Pass/Not Pass Удовлетвор./Неудовлетвор.
Measurement limit Предел измерения	20~600 mg/dL(1.1~33.3 mmol/L)	
Number of test measurements Число измерений	65	
Linear deviation Линейное отклонение	0.8837	PASS
Coefficient of variation Коэффициент вариации	2.979	PASS
Correlation coefficient Коэффициент корреляции	0.9962	PASS

Lot of product was tested and approved in accordance with Quality Control testing performance criteria

Signature / stamp / date

Owen 3/5/23

Frey 3/5/23



Foracare Suisse AG
Neugasse 55, 9000 St. Gallen
Switzerland
TEL: +41-71-220-1055
www.foracare.ch

DR. MATTHIAS JOHANN TILMANN

Візан 2016 05 200718